

08.02.2021

TAMILNADU MEDICAL SERVICES CORPORATION LTD.,

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF HIGH
FREQUENCY MOBILE X-RAY MACHINE AT VARIOUS VETERINARY INSTITUTIONS**

Tender No.602/XR/AHD/TNMSC/ENGG/2020, Dt.09.10.2020

Corrigendum

a) The following corrigendum are issued:-

Sl. No.	Tender document reference	Instead of	Read as
1.	Page No.52 Section V: Schedule of Requirements	Existing Text	Revised Text at Annexure I
2.	Page No.54 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine	<ul style="list-style-type: none">• Minimum power : 30 kW (300mA – 100kV @ 100 msec) or more• Minimum mAs : 300 or more• Minimum mA : 400 mA	High Frequency X-ray Generator with 30 KW output power or more with 40 to 125 KVP range and mA range of 125 to 300mA.
3.	Page No.54 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine X-ray tube Collimator	Large focus - 1.2 or less output at 75KW or more. Small Focus – 0.6mm or less with output 25KW or more. Tube with anode heat storage capacity 300 KHU or more.	Large focus - 1.3 mm or less output at 32KW or more. Small Focus – 0.6mm or less with output 11 KW or more. Tube with anode heat storage capacity 100 KHU or more.
4.	Page No.54 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine	Manual Multi leaf collimator with auto light shut-off. Halogen lamp with at least 100 watts bulb.	Manual two pairs of symmetric collimator with auto light shut-off. Light source with intensity of 150 LUX or more.

Sl. No.	Tender document reference	Instead of	Read as
5.	Page No.54 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine Operation Mode	mA Selection : from 25mA to 400mA	mA Selection : from 40mA to 400mA
6.	Page No.55 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine Mobile Stand	Should allow lowering of X-ray tube vertically to a minimum of 30 cm from the ground level.	Should allow lowering of X-ray tube vertically to a minimum distance required from x-ray sources to cassette/film is 60 cm from the ground level.
7.	Page No.55 Section VI: Technical Specification Specification for High Frequency Mobile X-ray Machine Power supply	Electric Power Consumption: 16 A maximum.	Power supply 1-phase 230 V (AC supply) 50/60 Hz \pm 10% with automatic compensation. 15 ampere wall socket.
8.	Page No.57 Section VI A: Qualification Criteria 3. Common to both manufacturer and bidder	a) The bidder should have a manufacturing capacity of at least 198 Nos. similar capacity machines per annum certified by licensing authority/ chartered accountant.	a) The bidder should have a manufacturing capacity of at least 117 Nos. similar capacity machines per annum certified by licensing authority/ chartered accountant.
9.	Page No.57 Section VI A: Qualification	b) The annual sales turnover for the bidder should be at	b) The annual sales turnover for the bidder should be at

Sl. No.	Tender document reference	Instead of	Read as
	Criteria 3. Common to both manufacturer and bidder	least Rs.96.00 lakhs in any one of the last three years (2017-2018 to 2019-2020), and it should be certified by the chartered accountants.	least Rs.113.60 lakhs in any one of the last three years (2017-2018 to 2019-2020), and it should be certified by the chartered accountants.
10.	Page No.57 Section VI A: Qualification Criteria 3. Common to both manufacturer and bidder	c) The bidder should have a cash/ credit limit of at Rs.24.00 lakhs with their bankers exclusively for this contract and a certificate for the above shall be furnished by the bankers.	c) The bidder should have a cash/ credit limit of at Rs.28.40 lakhs with their bankers exclusively for this contract and a certificate for the above shall be furnished by the bankers.
11.	Page No.59 Section VII: Bid Form and Price Schedule Price Schedule	Existing Text	Revised Text at Annexure II

a) The following clarification are furnished:-

Sl. No.	Tender document reference	Point raised by the bidder	Clarification furnished
1.	Page No.46 Section IV: Special Condition of Contract 10. Payment Payment for Goods and Services shall be made in Indian Rupees as follows: (i) No advance payment is payable.	Request to amend as upon receipt of Installation, verified Annexure & Delivery challan TNMSC will release 90% payment & help to release the Stock Entry certificate.	No Change. Hence, published terms prevails.

Sl. No.	Tender document reference	Point raised by the bidder	Clarification furnished
	(ii) 100% payment will be made against supply and installation of the equipments at the respective sites against certification along with necessary stock entry bills from the respective consignee.		

All other terms and conditions of the tender remain unaltered.

The above forms part of the bidding documents. The bidder shall attach the copy of this corrigendum duly signed by their authorized signatory, in their bid.

**Sd/-
General Manager (E)**

**SECTION – V - REIVSED
SCHEDULE OF REQUIREMENTS**

Sch. No.	Brief Description	Unit	Qty.	Bid security (Rs.)
1.	High Frequency Mobile X-ray Machine as per specification	Nos.	13	Rs.1,10,000/-

Delivery Schedule: -60 days from the date of purchase order and date of e-LORA permission whichever is later for the full quantity and for increased quantity the delivery period will be increased proportionately.

Place of Delivery:- The above equipment should be delivered at following places:-

1. Veterinary Polyclinics at Dindigul – 1 No.
2. Veterinary Polyclinics at Erode – 1 No.
3. Veterinary Polyclinics at Thanjavur – 1 No.
4. Veterinary Polyclinics at Thoothukudi – 1 No.
5. Veterinary Polyclinics at Tiruppur – 1 No.
6. Veterinary Polyclinics at Vellore – 1 No.
7. Up-gradation of the Clinican Centre at Gobichettipalayam in Erode District – 1 No.
8. Veterinary Hospital at Udumalaipettai in Tirupur District - 1 No.

Sl. No.	District	Upgradation Veterinary Hospital	Qty.
1.	Coimbatore	Kolarpatti	1 No.
2.	Dharmapuri	Karimangalam	1 No.
3.	Namakkal	Kuppandalpalayam	1 No.
4.	Tiruppur	Palladam	1 No.
5.	Tiruvarur	Thiruvizhimizhalai	1 No.

Note :

- 1) The rate contract should be valid for a period of 2 years from the date of acceptance of the tender and extendable for further period on mutual acceptance.
- 2) The quantity indicated is tentative and will vary as per the actual requirement during the rate contract period. The supplier should have no claim against the purchase on the actual quantity of orders placed during the rate contract period.
- 3) During the rate contract period, further quantity may be ordered for any other Govt. Medical Institutions at the same rate and terms & conditions.

ANNEXURE II

PRICE SCHEDULE - REVISED

1	2	3	4	5					6	7	8
				PRICE FOR EACH UNIT							
Sch No.	Item Description	Country of origin	Quantity & Unit	Ex-factory Ex-warehouse Ex-showroom off-the shelf for CIF (a)	Packing & forwarding (b)	Inland transport, Insurance and Incidental costs incidental to delivery (c)	Incidental services as listed in clause 7 of SCC (d)	Customs duty (e)	Unit price a+b+c+d+e	Total price per schedule for delivery at final destination (4 x 6)	GST – IGST/CGST/ SGST payable if contract is awarded
1.	High Frequency Mobile X-ray Machine as per specification		13 Nos.								

i. Unit price in (6) (Rs. in words)

1. High Frequency Mobile X-ray Machine as per specification with HSN code/rate of GST.....

a) Annual Maintenance Charges (labour only) per year /per unit for 7 years after 3 years free warranty maintenance period

1. High Frequency Mobile X-ray Machine as per specification with HSN code/rate of GST.....

b) Annual Maintenance Charges (Comprehensive) for 7 years / per year / per unit after free warranty maintenance period

1. High Frequency Mobile X-ray Machine as per specification with HSN code/rate of GST.....

Note:

- (a) In case of discrepancy between unit price and total price, the unit price shall prevail.
- (b) **This price schedule should be placed in separate sealed cover “Cover B”**
- (c) **GST applicable for Annual Maintenance Charges shall be indicated separately.**
- (d) **The bidder should indicate the HSN code of the equipment/ service and applicable GST rates.**
- (e) **The bidder should quote for the prices separately from column 5(a) to 5(e) and should not state “as inclusive”.**

Place:

Signature of Bidder.....

Date:

Name

Business Address