Dt. 14.05.2018

EVALUATION REPORT

Tender for fixing rate contract for supply and installation of Anesthesia Workstation to various Govt. Head Quarters Hospitals

Tender No. 262/ANS/DMS/TNMSC/ENGG/2018, Dt. 10.01.2018

---

a) Anesthesia Workstation (Type-I)

Sl. No	Name of the firm	Demo report	Commercial evaluation report	Reason for rejection
1	**M/s. Allied Medical Ltd.,** Model : Neptune plus with M747 Monitor Make : Meditec England	Non-responsive	Responsive	The quoted model does not meet the following points : Electronically controlled electrically driven ventilator should not require any driving gas. PEEP : 3-30 mm Hg instead of 0~20 mm Hg as per tender specification. Should be able to ventilate with atmospheric air, in case of missing gases.
2	**M/s. Madras Surgicals & Chemicals** Model : Fabius Plus Make : Draeger	**Responsive**	**Responsive**	**-----**
3	**M/s. Transhealthcare India Pvt. Ltd.,** Model : CWM302 with G3D Multiparameter Monitor Make : Nanjing Chenwei Medical Equipment Co. ltd., China	Non-responsive	Responsive	The quoted model does not meet the following points : Electronically controlled electrically driven ventilator should not require any driving gas. PEEP : 3-30 mm Hg instead of 0~20 mm Hg as per tender specification.
4	**M/s. BPL Medical Technologies Pvt. Ltd.,** Model : EFLO 6D Make :BPL	Non-responsive	Responsive	The quoted model does not meet the following points : Electronically controlled electrically driven ventilator

Sl. No	Name of the firm	Demo report	Commercial evaluation report	Reason for rejection
				should not require any driving gas. PEEP : 4-30 mm Hg instead of 0~20 mm Hg as per tender specification. Should be able to ventilate with atmospheric air, in case of missing gases.

b) Anesthesia Workstation (Type-II)

Sl. No	Name of the firm	Demo report	Commercial evaluation report	Reason for rejection
1	**M/s. Allied Medical Ltd.,** Model : Jupiter Plus with M747 Monitor Make : Allied Medical	**Responsive**	**Responsive**	**-----**
2	**M/s. Green Apple Medical Systems** Model : Wato EX-20 Make : Mindray	**Responsive**	**Responsive**	**-----**
3	**M/s. Hospitech** Model : Urjamint with Autoclavable Circle Absorber with Anesthesia Ventilator with Vapourizer (Isoflurane/Sevoflurane) with Monitor Make : Hospitech	**Responsive**	**Responsive**	**-----**
4	**M/s. Madras Surgicals & Chemicals** Model : Elite 915 Make : Avasarala	**Responsive**	**Responsive**	**-----**
5	**M/s. Transhealthcare India Pvt. Ltd.,** Model : CWM302 with G3D Multiparameter Monitor Make : Nanjing Chenwei Medical Equipment Co. ltd., China	**Responsive**	**Responsive**	**-----**
6	**M/s. BPL Medical Technologies Pvt. Ltd.,** Model : EFLO 6D Make :BPL	**Responsive**	**Responsive**	**-----**