Dt.14.11.2018

EVALUATION REPORT

Tender for Fixing Rate Contract for supply of Laryngoscope Neonate to various Govt. Institutions

Tender no.361/SCOPE/NHM/TNMSC/ENGG/2018, Dt.04.09.2018.

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
Equipment: Laryngoscope Neonate – 20 nos				
1	M/s. Alan Medical and Laboratory Products, Make: Alan	**Responsive**	**Responsive**	-----
2	M/s. Medifocus India Pvt. Ltd, Make: Karl Storz	Non-Responsive	Non-Responsive	• Against the technical point: 2 x straight depressor, Miller type No.0 & 00 with fibre optic system – only Miller type No. 0 is available. • Did not furnish the end user certificate with supporting document like PO, Installation report for similar system to meet the qualification criteria. The documents furnished are only for the manufacturer.