05.12.2019

Evaluation Report

Tender No. 448/SCOPE/DMS-ESI/TNMSC/ENGG/2019, Dt. 04.03.2019

Tender for supply and installation of Basic Diagnostic Arthroscopy set to ESI hospitals at Ayanavaram, Madurai and Salem

Sch. No. (1A) : *Basic Diagnostic Arthroscopy Set with inbuilt full HD Video and Image Recording System*

Sch. No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1A	M/s. Vishal Surgical Equipment Co. Pvt. Ltd. Model : Telescope - Hopkins 30° (28731BWA) Drill Cut - X Arthro Shaver Hand piece (28200DX) Camera - Karl Storz - SCB (TC200EN) 27" Monitor (TM220) Trolley (Indian Make) Make : Karl Storz, Germany	Non Responsive	Responsive	**Demo Report** a) The quoted model is not meeting the technical specification of microprocessor control console should accept three-hand pieces and should run two-hand pieces with two different footswitches at a time simultaneously. b) The quoted model is not user friendly to operate, hence not recommended.

Sch. No. (1B) : *Basic Diagnostic Arthroscopy Set with standalone full HD Video and Image Recording System*

Sch. No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1B	M/s. Indian Orthocare Surgicals Model : Stryker 4 mm 30 deg Ideal Eyes Arthroscope (402104030) Stryker KIT, Formula Shaver Hand Piece with Button (375704500E) Stryker 1488 Camera Console & Camera Head Visionpro 26' LED Display IOS Endoscopy Trolley with Flexible Arm Make : Stryker, USA	Responsive	Responsive	----------