EVALUATION REPORT

Tender no.526/OTT/DME-RGGGH/TNMSC/ENGG/2019, Dt.09.09.2019

Tender for supply and installation of Fully Motorized C-Arm Compatible OT Table to Institute of Vascular Surgery, Rajiv Gandhi Govt. General Hospital, Chennai

Sch. No	Name of the Equipment and Bidder	Responsiveness of bidder		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Solokrafts Industries, Make: Mindray, Model: Hybase 6100	Non-Responsive	Not evaluated	Not meeting the following technical points: 1. Against the technical specification of "Table top length should be at least 220 cm with imaging length of 160 cm or more and width of the table 600 mm or more" – offered width of the table 590 mm. 2. Against the technical specification of "Table should have motorized longitudinal travel 90 cm and lateral travel of 25 cm or more" – lateral travel not available.