Date: 12.02.2020

TENDER FOR SUPPLY AND INSTALLATION OF INFUSION PUMP AND SYRINGE PUMP FOR MEDICAL COLLEGE HOSPITALS AND DISTRICT HEAD QUARTERS HOSPITAL AT VARIOUS DISTRICTS OF TAMIL NADU

PKG14/LCB/TNUHP/TNMSC/ENGG/2019, DT:25.10.2019

Result of Evaluation

Item Name: Sch.no.1: INFUSION PUMP– 8 Nos.

Sch. No.	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation	Commercial Evaluation	
1.	M/s. Akas Medical, Chennai	*Make*: Akas, India *Model*: Infumax	Responsive	Responsive	------
	M/s. Medifocus India Private Ltd., Chennai	*Make*: BBraun, USA *Model*: Infusomat	Responsive	Responsive	------
	M/s. South India Surgical Co. Ltd., Chennai	*Option 1*: *Make*: Enmind, China *Model*: EN-V7 Smart	Responsive	Responsive	------
		Option 2: *Make*: JMS, Japan *Model*: OT-701	Non-Responsive	Non-Responsive	**Technical Evaluation:** Item Not produced for demo. Hence, non-responsive. **Commercial Evaluation:** 1. Technical Compliance statement not furnished. 2. Annual Manufacturing Capacity not furnished. 3. Email from M/s SISCO dt 26.12.2019 stating the Infusion pump in option 2 quoted model doesn't match the tender specifications. Hence, Option 1 alone is considered for evaluation.

Item name: Sch.no.2: SYRINGE PUMP- 8 Nos.

Sch. No.	Name of the Bidders	Equipment Details	Responsiveness		Reason for Rejection
			Technical Evaluation	Commercial Evaluation	
2.	M/s. Akas Medical, Chennai	*Make*: Akas, India *Model*: Terumax RX	Responsive	Responsive	------
	M/s. Medifocus India Private Ltd., Chennai	*Make*: BBraun, USA *Model*: Compact Plus	Responsive	Responsive	------
	M/s. South India Surgical Co. Ltd., Chennai	*Option 1*: Make: Enmind, China Model: EN-S7 Smart	Responsive	Responsive	------
		Option 2: Make: JMS, Japan Model: SP-500	Non-Responsive	Non-Responsive	**Technical Evaluation:** Item Not produced for demo. Hence, non-responsive. **Commercial Evaluation:** 1. Technical Compliance statement not furnished. 2. Annual Manufacturing Capacity not furnished. 3. Email from M/s SISCO dt 26.12.2019 stating the Infusion pump in option 2 quoted model doesn't match the tender specifications. Hence, Option 1 alone is considered for evaluation.

Sd/-

General Manager (E)