19.03.2020

SUPPLY AND INSTALLATION OF VENTILATOR FOR MEDICAL COLLEGE HOSPITALS AT VARIOUS DISTRICTS OF TAMIL NADU

TENDER NO. PKG12/LCB/TNUHP/TNMSC/ENGG/2019, DT:25.10.2019

Result of Evaluation

Ventilator– 13 Nos.

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Madras Surgicals & Chemicals	**Model**: Savina 300 Select **Make**: Drager	Non-Responsive	Responsive	Does not comply to the following technical specifications: 1. Shall be suitable for continuous use in ICU for therapeutic application on hospital Air & Oxygen supply. 2. Suitable standalone external air compressor should also be provided and it shall be mounted on ventilator trolley itself. 3. The external air compressor provided by the manufacturer shall be of medical grade, noise level not more than 50dB, weight not more than 30 Kg. Shall have automatic change over facility when central gas supply fails. The external air compressor shall be capable of delivering 30 LPM continuous flow dry air, reservoir capacity of 1.25L & shall be supplied with inlet moisture filter for ventilators. The external air compressor shall have FDA 510(K) or CE marking. 4. Should provide pressure and flow triggering options. 5. Shall have following setting parameters:-Flow Pattern: Square & Decelerating 6. Shall have display of waning parameters like: Expiratory time constant

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
2.	M/s. Schiller Healthcare India Pvt.Ltd.	**Model**: Graphnet TS **Make**: Tecme (Schiller Neumovent)	Responsive	Responsive	NIL
3.	M/s. Indian Medical Systems and Technologies	**Model**: Servo i **Make**: Getinge	Responsive	Responsive	NIL
4.	M/s. Trans Healthcare India Pvt. Ltd.	**Model**: iX5 **Make**: Vyaire Medical	Responsive	Responsive	NIL