

**TENDER FOR SUPPLY AND INSTALLATION OF VESSEL SEALING
MACHINE AT VARIOUS VETERINARY INSTITUTIONS**

TENDER NO. 605/DIAT/AHD/TNMSC/ENGG/2020, DT: 09.10.2020

Vessel Sealing Machine – 8 No.

Sl. No.	Bidders Quoted Model/ Make	Responsiveness		Reason for rejection / Remarks
		Technical evaluation report	Commercial evaluation report	
1.	M/s. JBS Mediitec India Pvt. Ltd. Model: Combi max MBxVxMake: Alan Electronic Systems Pvt. Ltd.	Responsive	Responsive	-----
2.	M/s. XcellLance Medical Technologies Pvt. Ltd. Mumbai. Model: Vista, Make: Own	Responsive	Responsive	-----
3.	M/s. The Advanced Medical Systems Inc. Option-I-Excel-400 Option-II-Excel-NHP Make: - Alsa Apparecchi Medicali S R I – Italy.	Responsive	Responsive	-----
4.	M/s. Suraj Meditec Agency. Make: Tanvi Electromedical Appliances. Model: Tanvi Med Sealer with touch screen display	Non - Responsive	Responsive	Based on the specialist report: - "The demonstrated model was loud sound whenever we use the vessel sealer and the coagulation unit". Hence not recommended.
5.	M/s. Thirupathi Enterprises Option-I-Easemed-Manoseal Prime. Option-II-Easemed-Excel Sealer. Make:- M/s. Easemed Electrosurgical Pvt. Ltd.	Non - Responsive	Non - Responsive	Not demonstrated; hence quality could not be ascertained. Hence not-recommended.