

Dt.16.04.2021

## EVALUATION REPORT

Tender no.576/CEN/RC/TNMSC/ENGG/2020, Dt.17.06.2020

**Tender for Fixing Rate contract for supply of Centrifuge, Refrigerated Centrifuge and Centrifuge-Sero Fuge to various Govt. Institutions in the State**

**Sch.no.1 Centrifuge – 10 nos. tentative quantity**

Sl. No	Name of the Equipment and Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<b>M/s. Agile Lifescience Technologies India Pvt. Ltd,</b> Make: Andreas Hettich, Germany, Model: Rotofix 32A	<b>Responsive</b>	<b>Responsive</b>	-----
2	<b>M/s. Alan Medical &amp; Laboratory Products,</b> Make: Remi, Model: Neya 10, S4-175. B-175, 17-5/7, A8-50, I7-10, RE 15-5/7L, A32-15, I12-2, I1-100	Non-responsive	Responsive	As per the technical literature quoted make does not meeting the following points. <ul style="list-style-type: none"> <li>• The centrifuge should be capable of accepting a 12 slide cyto rotor, and must be clearly mentioned with details of the cyto rotor system – <u>Mentioned as <b>not available.</b></u></li> <li>• Capacity – Accommodate from 133 tubes of 0.5ml to 4 tubes of 100 ml. Adapters should be available for tubes of volume range between 0.5 ml to 100 ml including vacutainers. Clinical labs have many tubes for processing and versatile centrifuge is preferable – <u>mentioned as capacity accommodate from 2ml to 4x100ml &amp; adapters provided for 2ml to 100ml including vacutainer - <b>The number of tubes and capacity required does not matching.</b></u></li> </ul>
3	<b>M/s. Inkarp Instruments Pvt. Ltd,</b> Make: Andreas Hettich, Germany, Model: Rotofix 32A	Non-responsive on bid security		

### Sch.no.2 Refrigerated Centrifuge – 10 nos. tentative quantity

Sl. No	Name of the Equipment and Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<b>M/s. Agile Lifescience Technologies India Pvt. Ltd,</b> Make: Andreas Hettich, Germany, Model: Rotosilents 630RS	<b>Responsive</b>	<b>Responsive</b>	-----
2	<b>M/s. Alan Medical &amp; Laboratory Products,</b> Make: Labtop, Model: LBC-12	Non-responsive	Responsive	<ul style="list-style-type: none"> <li>Against the specification "Should be USFDA or European CE certified as class II A medical device" – Furnished CE certificate of Regal Register UK Ltd, which is <b><u>not a European CE notified body.</u></b></li> </ul>
3	<b>M/s. Inexus Biotech Pvt. Ltd,</b> Make: Himac, Japan, Model: CR 7N	Non-responsive	Not evaluated	<p>As per the technical literature quoted make does not meeting the following points.</p> <ul style="list-style-type: none"> <li>Max. Capacity: 12,000 ml should accommodate twelve 350ml and/or 450 ml single, double, triple, quadruple / quintuple blood bags with SAGM bag and empty satellite bags with in line filter system – offered <b><u>only 6000 ml.</u></b></li> <li>Dimensions (H x W x D) mm: 973 x 813 x 1015 mm, offered <b><u>730 x 845 x 928 mm (smaller size).</u></b></li> <li>Should be USFDA or European CE certified as class II A medical device – Mentioned as CE but <b><u>certificate not provided.</u></b></li> </ul>
4	<b>M/s. Inkarp Instruments Pvt. Ltd,</b> Make: Kubota, Japan, Model: 9942	Non-responsive on bid security		

### Sch.no.3 Centrifuge-Sero Fuge – 10 nos. tentative quantity

Sl. No	Name of the Equipment and Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	<b>M/s. Alan Medical &amp; Laboratory Products,</b> Make: Labtop, Model: LLC-4	<b>Responsive</b>	<b>Responsive</b>	-----