

**Tender for Fixing Rate contract for supply and Installation of Deep Freezer -40 deg C to various Govt. Institutions**

**Tender No. 577/DF/RC/TNMSC/ENGG/2020 Dt.17.06.2020**

**RESULTS OF EVALUATION**

**Deep Freezer -40 deg C:-**

<b>S. No</b>	<b>Name of the bidder</b>	<b>Technical Evaluation report</b>	<b>Commercial evaluation report</b>	<b>Reason for Rejection</b>
<b>1</b>	<b>M/s. Sub Zero Instruments</b> Model: Ultra low temperature Make: Own	Not evaluated	Not evaluated	Not evaluated due to non – responsiveness on bid security
<b>2</b>	<b>M/s. Sun Biomedical Pvt Ltd</b> Model: DF-40U 0 400L Make: Terumopenpol	Responsive	Responsive	-----
<b>3</b>	<b>M/s. Span Surgical Co</b> Model: M-Freeze 600 Make: MVTEX Science Industries	Non – Responsive	Non – Responsive	<b>In technical points:</b> Did not participated in demonstration <b>In commercial points</b> Not submitted Documentary evidence for satisfactory performance of the item quoted (client’s certificates)
<b>4</b>	<b>M/s. Solokrafts Industries</b> Model: EQF- 600 Make: Elektrocraft	Non – Responsive	Non – Responsive	<b>In technical points:</b> Did not participated in demonstration <b>In commercial points</b> 1. Not submitted Documentary evidence for satisfactory performance of the item quoted (client’s certificates) 2. Not furnished manufacturing capacity of at least 180 Nos. Similar capacity machines per annum certified by licensing authority / chartered accountant. 3. Not agreed for Delivery Schedule as per bid document (30 days from the date of receipt of the PO) 4. Not confirmed Warranty is as per bid document (3 Years from the acceptance of goods)

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
5	<b>M/s. Alan Medical and Laboratory Products</b> Model: LBF-600 Make: LABTOP	Non – Responsive	Non – Responsive	<b>In technical points:</b> Did not participated in demonstration <b>In commercial points</b> 1. Not submitted Documentary evidence for satisfactory performance of the item quoted (client's certificates) 2. Not furnished manufacturing capacity of at least 180 Nos. Similar capacity machines per annum certified by licensing authority / chartered accountant. 3. Not agreed for Delivery Schedule as per bid document (30 days from the date of receipt of the PO) 4. Not agreed Warranty is as per bid document (3 Years from the acceptance of goods) 5. Not furnished Spares parts availability for 10 years after sales service facilities in India 6. Not agreed for Payment Terms as per bid document.
6	<b>M/s. M.C Dalal and Co</b> Model: MCD 040 Make: Own	Non – Responsive	Responsive	<b>In technical points:</b> 1. Did not attain internal control temperature range of -45. (only reached -38) 2. Did not attain hold over time of minimum of 2 hours (only reached 25 minutes)
7	M/s. Asean Scientific Express <b>Option 1</b> <b>Model:</b> DF-40L400 <b>Make:</b> Own	Non – Responsive	Responsive	<b>In technical points:</b> 1. Did not attain internal control temperature range

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
	<b>Option 2</b> <b>Model:</b> DW-40L626 <b>Make:</b> Own			of -45. (only reached -38) 2. Did not attain hold over time of minimum of 2 hours (only reached 35 minutes)
8	M/s. Cryo Scientific Systems Pvt Ltd Model: URC-V-625-2 Make: Own	Non – Responsive	Responsive	<b>In technical points:</b> 1. Did not attain internal control temperature range of -45. (only reached -40) 2. Did not attain hold over time of minimum of 2 hours (only reached 95 minutes)