

30.04.2021

**INTERNATIONAL COMPETITIVE BIDDING (ICB) FOR SUPPLY AND
INSTALLATION OF OPERATION THEATRE EQUIPMENT FOR MEDICAL
COLLEGE HOSPITALS AND DISTRICT HEADQUARTERS HOSPITALS**

TENDER NO. PKG6/ICB/TNUHP/ JICA/C2/OPE-I/TNMSC/ENGG/2020

dated 04.02.2020

RESULT OF EVALUATION

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
Lot/contract No.1: Multipara monitor				
1.	M/s MM Medical Systems, Make: Nihon Kohden, Japan Model: SVM 7621K	Responsive	Responsive	-----
2.	M/s. Wipro GE Healthcare Private Limited Make: GE Model: B125P	Responsive	Responsive	-----
3.	M/s Green Apple Medical Systems Make: Mindray Model: uMEC12	Responsive	Responsive	-----
4.	M/s Trans Health Care India Pvt. Ltd., Make: General Meditec, China Model: G3D	Responsive	Responsive	-----
Lot/contract No.2: Multipara monitor with IBP				
1.	M/s MM Medical Systems, Make: Nihon Kohden, Japan Model: PVM 4763	Responsive	Responsive	-----
2.	M/s. Wipro GE Healthcare Private Limited Make: GE Model: B105M	Responsive	Responsive	-----

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
3.	M/s Green Apple Medical Systems Make: Mindray Model: uMEC12	Responsive	Responsive	-----
4.	M/s Trans Health Care India Pvt. Ltd., Make: General Meditec, China Model: G3D	Responsive	Responsive	-----
Lot/contract No.3: Multipara monitor with EtCO2				
1.	M/s MM Medical Systems, Make: Nihon Kohden, Japan Model: Lifescope VS BSM 3562	Responsive	Responsive	-----
2.	M/s. Wipro GE Healthcare Private Limited Make: GE Model: B125M	Responsive	Responsive	-----
3.	M/s Green Apple Medical Systems Make: Mindray Model: uMEC12	Responsive	Responsive	-----
4.	M/s Trans Health Care India Pvt. Ltd., Make: General Meditec, China Model: G3D	Responsive	Responsive	-----
Lot/contract No.4: Defibrillator				
1	M/s MM Medical Systems, Make: Nihon Kohden, Japan Model: Cardilife TEC 5631K	Responsive	Responsive	-----
2	M/s Green Apple Medical Systems Make: Mindray Model: Beneheart D3	Responsive	Responsive	-----

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
3	M/s Trans Health Care India Pvt. Ltd., Make: General Meditec, China Model: DM7000	Non -Responsive	Responsive	As per specialist remarks, the quoted model does not meet the following technical specification: 'Shall have Real time CPR Help. (Displaying compression Depth & Rate)'.
Lot/contract No.5: Anesthesia Workstation with AGM				
1	M/s Draeger India Pvt. Ltd., Make: Drager, Germany Model: Perseus A500	Responsive	Responsive	-----
2	M/s. Wipro GE Healthcare Private Limited Make: GE Model: Aisys CS2	Non -Responsive	Responsive	As per specialist remarks, the firm didn't furnish any documentary proof to justify the point "Patient ventilation and delivery of anesthesia shall function uninterrupted during that time and it should be able to ventilate with atmospheric air, in case of emergency oxygen failure/ any other technology as patient safety measure."
3	M/s Green Apple Medical Systems Make: Mindray, China Model: A9	Non -Responsive	Responsive	As per specialist remarks, the firm didn't furnish any documentary proof to justify the point "Patient ventilation and delivery of anesthesia shall function uninterrupted

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				during that time and it should be able to ventilate with atmospheric air, in case of emergency oxygen failure/ any other technology as patient safety measure."
Lot/contract No.6: Anesthesia Workstation				
1	M/s Madras Surgicals and Chemicals Make: Drager, Germany Model: Fabius Plus with Vapor 2000	Responsive	Responsive	-----
2	M/s. Wipro GE Healthcare Private Limited Make: GE Model: 9100c NXT	Non -Responsive	Responsive	As per specialist remarks, the firm didn't furnish any documentary proof to justify the point "Should be able to ventilate with atmospheric air in case of missing gas/ or any other technology for safety of the patients."
3	M/s Green Apple Medical Systems Make: Mindray, China Model: Wato EX-30	Non -Responsive	Responsive	As per specialist remarks, the documentary evidence furnished by the firm to justify the point "Should be able to ventilate with atmospheric air in case of missing gas/ or any other technology for safety of the patients", was not furnished from any reputed Medical

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				College. The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is already occupied by many equipment being used by Anesthesia and Surgical Specialties. So at this stage, models with such arrangement are not acceptable.
4	M/s BPL Medical Technologies Pvt. Ltd., Make: BPL, India Model: Eflo-6D	Non -Responsive	Responsive	As per specialist remarks, the documentary evidence furnished by the firm to justify the point "Should be able to ventilate with atmospheric air in case of missing gas/ or any other technology for safety of the patients", was not furnished from any reputed Medical College. The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				already occupied by many equipment being used by Anesthesia and Surgical Specialties. So at this stage, models with such arrangement are not acceptable.
5	M/s Trans Health Care India Pvt. Ltd., Make: Shenzhen Comen Medical Instruments, China Model: AX-400	Non -Responsive	Responsive	As per specialist remarks, the documentary evidence furnished by the firm to justify the point "Should be able to ventilate with atmospheric air in case of missing gas/ or any other technology for safety of the patients", was not furnished from any reputed Medical College. The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is already occupied by many equipment being used by Anesthesia and Surgical Specialties. So at this stage, models with such arrangement are not acceptable.

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
Lot/contract No.7: Operation Theatre Table with Ortho Attachments, C-ARM compatible				
1	M/s Confident Dental Equipments Pvt.Ltd., Make: Confident, India Model: EHOT-012	Substantially Responsive	Responsive	-----
2	M/s Staan Biomed Engineering Pvt. Ltd., Make: Staan, India Model: Glory	Responsive	Responsive	-----
3	M/s South India Surgical Co. Ltd., Make: Galaxy, India Model: Asira Series	Non - Responsive	Non - Responsive	<u>Technical point:</u> As per specialist remarks, the quoted model does not meet the following technical specification: 'Maximum patient weight: 360 kgs', against which they have provided compliance for 300 kg. <u>Commercial terms:</u> The Manufacturer does not meet the minimum average annual turnover.
4	M/s Green Apple Medical Systems Make: Mindray, China Model: Hybase 6100 Plus	Non - Responsive	Responsive	As per specialist remarks, the quoted model has only locking via remote which does not comply to the tender specification, 'locking via foot pedal' and also zero position

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				only via remote control which does not comply to the tender specification 'zero position should be operated from the remote hand set as well as the override panel'
Lot/contract No.8: Operation Theatre Table with General surgery attachments				
1	M/s Confident Dental Equipments Pvt.Ltd., Make: Confident, India Model: EHOT-012	Responsive	Responsive	-----
2	M/s Staan Biomed Engineering Pvt. Ltd., Make: Staan, India Model: Glory	Responsive	Responsive	-----
3	M/s South India Surgical Co. Ltd., Make: Galaxy, India Model: Asira Series	Non - Responsive	Non - Responsive	As per specialist remarks, the quoted model does not meet the following tender specification: i) Dia of swivel castors with breaks: 125 mm (minimum) (with a $\pm 10\%$ permitted deviation) against which the compliance given by the bidder is 100 mm. ii) Longitudinal displacement: more than 350 mm (with a $\pm 10\%$ permitted deviation) against

Sl. No	Bidders Quoted (with make and model)	Responsiveness		Reason for rejection / Remarks
		Technical Responsiveness	Commercial Responsiveness	
				<p>which the compliance given by the bidder is 300 mm.</p> <p>iii) Should have fifth wheel for 360 deg table rotation, against which the bidder has stated No.</p>
4	<p>M/s Green Apple Medical Systems</p> <p>Make: Mindray, China</p> <p>Model: Hybase 6100 Plus</p>	Non - Responsive	Responsive	<p>As per specialist remarks, the quoted model does not meet the following tender specification:</p> <p>i) Zero position should be operated from the remote hand set as well as the override panel against which the bidder has stated that zero position is available only on the remote control.</p> <p>ii) Dia of swivel castors with breaks: 125 mm (minimum) (with a $\pm 10\%$ permitted deviation) against which the compliance given by the bidder is 75 mm.</p>