

Dt.05.05.2021

EVALUATION REPORT

Tender for fixing rate contract for supply and installation of ECG Machine Computerized with accessories and consumables to various Government Institutions in the state.

Tender No.595/ECG/RC/TNMSC/ENGG/2020, Dt. 21.08.2020.

ECG Machine Computerized with accessories and consumables				
Sl. No	Name of the bidders	Demo report	Commercial evaluation report	Reason for rejection
1	M/s. Surya Surgical Industries Model : Cardiart 9108D Make : BPL	Responsive	Responsive	----
2	M/s. Madras Surgicals & Chemicals Model : Cardio 7 Make : Bionet	Responsive	Responsive	----
3	M/s. Bionic Biomedical Systems Model : E70 Make : Biolight, P.R. China	Responsive	Responsive	----
4	M/s. AMS Solutions Inc., Model : ECG 712 Series Make : NIDEK Medical India Pvt. Ltd.,	Non-responsive	Not evaluated due to non-responsiveness in technical evaluation	The quoted model has 12 bit A/D converter against the tender requirement of 24 bit A/D converter.
5	M/s. MM Medical Systems Model : CARDIFAX M ECG 2350 Make : Nihon Kohden, Japan	Responsive	Responsive	----
6	M/s. Sabisu Model : CM 1200B, Make : Comen Model : CM1200A, make : Comen	Responsive	Responsive	----