

EVALUATION REPORT

**Tender for fixing rate contract for supply and installation of
Anesthesia Workstation (Type-I & II) to various Government
Institutions in the state.**

Tender No.612/ANS/RC/TNMSC/ENGG/2020, Dt. 09.10.2020.

a) Anesthesia Workstation (Type-I)

Sl. No	Name of the bidders	Demo report	Commercial evaluation report	Reason for rejection
1	M/s. Transhealthcare India Pvt. Ltd., Model : AX-400 Make : Shenzhen Comen Medical Instruments Co. Ltd., P.R. China	Non-responsive	Responsive	The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is already occupied by many equipments being used by Anesthesia and Surgical Specialties.
2	M/s. BPL Medical technologies Pvt. Ltd., Model : EFLO 6D with Ultima Prime with dual IBP & sidestream EtCO2 Make : BPL	Non-responsive	Responsive	The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is already occupied by many equipments being used by Anesthesia and Surgical Specialties.
3	M/s. Madras Surgicals & Chemicals Model : Fabius Plus Make : Draeger	Responsive	Responsive	-----
4	M/s. Solokrafts Industries Model : Wato EX-30 with UMEC 12 Make : Mindray	Non-responsive	Responsive	The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space in operation theatres which is already occupied by many equipments being used by Anesthesia and Surgical Specialties.
5	M/s. Allied Medical Ltd., Model : Neptune Plus with M747 Monitor Make : Allied Meditec	Non-responsive	Responsive	The arrangement for ventilation with atmospheric air in case of missing gases, were found to be external to the Anesthesia machine. This will occupy more space

Sl. No	Name of the bidders	Demo report	Commercial evaluation report	Reason for rejection
				in operation theatres which is already occupied by many equipments being used by Anesthesia and Surgical Specialties.

b) Anesthesia Workstation (Type-II)

Sl. No	Name of the firm	Demo report	Commercial evaluation report	Reason for rejection
1	M/s. Transhealthcare India Pvt. Ltd., Model : AX-400 Make : Shenzhen Comen Medical Instruments Co. Ltd., P.R. China	Responsive	Responsive	----
2	M/s. BPL Medical technologies Pvt. Ltd., Option-I Model : EFLO 6D with Ultima Prime with dual IBP & AGM Make : BPL Option-II Model : Penlon Prima 320 with BPL Ultima Prime with dual IBP & AGM Make : Penlon & BPL	Responsive	Responsive	----
3	M/s. Solokrafts Industries Option-I Model : Wato EX-30 with UMEC 12 Make : Mindray Option-II Model : Wato EX-30 with e-PM12 Make : Mindray	Responsive	Responsive	----
4	M/s. Allied Medical Ltd., Model : Neptune Plus with M747 Monitor Make : Allied Meditec	Responsive	Responsive	----
5	M/s. AMS Solutions Inc Model : VIVID JS-400 Make : AMS	Non-Responsive	Responsive	Not demonstrated properly. Technical person not able to explain the feature. CE certificate not furnished.