

EVALUATION REPORT

Tender for fixing rate contract for supply and installation of Boyles apparatus (Type-I & II) to various Government Institutions in the state

Tender No.E625/ANS/RC/TNMSC/ENGG/2020, Dt. 17.12.2020.

a) Boyles apparatus (3 gas system)

Sl. No	Name of the bidders	Demo report	Commercial evaluation report	Reason for rejection
1	M/s. Avasarala Technologies Ltd., Model : Elite 615 Make : Avasarala	Responsive	Responsive	----
2	M/s. Suraj Meditec Agency Model : SMA 5 Make : Suraj Meditec	Responsive	Responsive	----
3	M/s. Transhealthcare India Pvt. Ltd., Model : Allied Jupiter 500 Make : Allied Medical	Responsive	Responsive	----

b) Boyles apparatus (2 gas system)

Sl. No	Name of the firm	Demo report	Commercial evaluation report	Reason for rejection
1	M/s. Avasarala Technologies Ltd., Model : Elite 415 Make : Avasarala	Responsive	Responsive	----
2	M/s. Suraj Meditec Agency Model : SMA 2 Make : Suraj Meditec	Responsive	Responsive	----
3	M/s. Transhealthcare India Pvt. Ltd., Model : Allied Jupiter 200 Make : Allied Medical	Responsive	Responsive	----