

TENDER FOR SUPPLY AND INSTALLATION OF NITRIC OXIDE DELIVERY

SYSTEM TO VARIOUS GOVT. INSTITUTIONS

TENDER NO. 618/NO/NHM/TNMSC/ENGG/2020, DT. 21.10.2020

RESULTS OF EVALUATION AND FINALLY SHORTLISTED BIDDERS

A: NITRIC OXIDE DELIVERY SYSTEM

Sl.No	Name of the firm	Technical evaluation report	Commercial evaluation report	Reason for rejection
1.	M/s. Shantech Healthcare Model: GeaNOX Make: M/s. Sicgilsol India Pvt Ltd	Responsive	Responsive	-----
2.	M/s. Peso Medicare Pvt Ltd Model: NOXtec 2000 Make: M/s. Ingenieria Y Tecnicas Clinicas, S.A	Non Responsive	Non Responsive	As per the demonstration report, the model quoted by the bidder does not meet with the following points, hence not recommended: a. Calibrating gas for NO and NO2 in a single cylinder (Two cylinders not available) against the tender requirement of two cylinders. b. Facility to change to the second cylinder by a simple turn of a valve is not available. The regulator has to be manually removed and fixed to the second cylinder against the tender requirement of facility to change to the second cylinder by a simple turn of a valve. Also not furnished end user certificate with supporting documents viz PO and installation report for the supplies made by their Manufacturer with 1 yr satisfactory performance along with the affidavit for the supplies made, Manufacturing capacity certificate and Cash credit limit certificate.