

**Tender for Fixing Rate Contract for supply and Installation of Lab Freezer
-80 deg C to various Govt. Institutions**

Tender No. 611/DF/RC/TNMSC/ENGG/2020 dt.07.01.2021

RESULT OF EVALUATION

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
1	M/s. Yorco Sales Pvt Ltd Model: YSI-273Ex Make: Own	Non – Responsive	Non – Responsive	<p>In technical Points:</p> <p>1. Did not participate in the demo and hence not recommended.</p> <p>In Commercial points:</p> <p>1. Not furnished the documentary evidence for satisfactory performance of the item quoted.</p> <p>2. Not furnished Manufacturing capacity of atleast 180 Nos. certified by licensing authority.</p> <p>3. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate.</p>
2	M/s. Sun Biomedical Pvt Ltd Model: DF-80 400 L Make: Terumopenpol	Responsive	Responsive	-----
3	M/s. Asean Scientific Express Model: DF – 80L 350 Make: Own	Non – Responsive	Responsive	<p>In technical points:</p> <p>1. Has no non-pull out stainless trays (Point No.2(b))</p> <p>2. No audio visual electronic alarm system independent of power supply. (No power failure indication, no high low temperature alarm/ indicators) (Point – Alarm (a))</p> <p>3. No servo stabilizer (Point No.4 (b))</p> <p>4. Also not having IS-13450 for electrical safety standards (Point No. 5 (b)) and hence not recommended.</p>
4a	M/s. Cryo Scientific Systems Pvt. Ltd Option 1 Model: URC-V-325-4 Make: Own	Non – Responsive	Non – Responsive	<p>In technical points:</p> <p>1. No audio visual electronic alarm system independent of power supply. (No power failure indication, no high low</p>

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
				<p>temperature alarm/ indicators) (Point – Alarm (a)) 2. Not having external graph recorder. (Point. 4 (d)). 3. Not having IS-13450 for electrical safety standards (Point No. 5 (b)) and hence not recommended</p> <p>In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted. 2. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate.</p>
4b	<p>M/s. Cryo Scientific Systems Pvt. Ltd Option 2 Model: URC-V-380-4</p>	Non – Responsive	Non – Responsive	<p>In technical points: 1. Did not participate in the demo and hence not recommended.</p> <p>In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted. 2. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate.</p>
4c	<p>M/s. Cryo Scientific Systems Pvt. Ltd Option 3 Model: URC-V-500-4 Make: Own</p>	Non – Responsive	Non – Responsive	<p>In technical points: 1. Did not participate in the demo and hence not recommended.</p> <p>In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted. 2. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate.</p>

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
5	M/s. Span Surgical Co Model: M – Freezer 800 Make: MV Tex	Non – Responsive	Non – Responsive	In technical Points: 1. Did not participate in the demo and hence not recommended. In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted. 2. Not furnished Performance statement X. 3. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate. 4. Not furnished cash/ credit limit of at Rs.12.25 lakhs from the banker
6	M/s. Alan Medical and Laboratory Products Model: LUPF – 350 Make: Labtop	Non – Responsive	Non – Responsive	In technical Points: 1. No non-pull stainless trays (Point No.2(b)) 2. No audio visual electronic alarm system independent of power supply. (No power failure indication, no high low temperature alarm/ indicators) (Point – Alarm (a)) 3. No servo stabilizer and no battery backup available (Point No.4 (b)) 4. No battery backup available (Point No. 4 (d)). 5. Also not having IS-13450 (Point No. 5 (b)) for electrical safety standards and hence not recommended In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted.
7	M/s. Eppendorf India Pvt Ltd Model: Premium U410 Upright Freezer Make: Own	Non – Responsive	Non – Responsive	In technical Points: 1. Did not participate in the demo and hence not recommended. In Commercial points: 1. Not furnished the documentary evidence for satisfactory performance of the item quoted.

S. No	Name of the bidder	Technical Evaluation report	Commercial evaluation report	Reason for Rejection
				<ol style="list-style-type: none"><li data-bbox="954 248 1524 360">2. Not furnished Manufacturing capacity of atleast 180 Nos. certified by licensing authority.<li data-bbox="954 371 1524 564">3. Not furnished affidavit conforming that the performance statement given is correct along with the copy of purchase order, copy of invoices and end user certificate.