

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION
OF MODULAR PATIENT MONITOR AND HIGH END MULTICHANNEL
MONITOR TO VARIOUS GOVT. INSTITUTIONS**

TENDER NO.593/MON/RC/TNMSC/ENGG/2020, DT.21.08.2020

Result of Evaluation

Modular Patient Monitor – 5 nos.

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1.	M/s. Trans Health Care India Pvt. Ltd.	Sino-Hero (Shenzhen) Bio Medical Electronics Co. Ltd., China 0/ ACUITSIGN M8	Responsive	Responsive	_____
2.	M/s. Solokrafts Industries.	Benevision N15+N1	Responsive	Responsive	_____
3.	M/s. Sri Sai Ekdant Surgicals.	Philips MX 550	Responsive	Responsive	_____
4.	M/s. M M Medicals Systems.	Lifescop TR BSM 6701	Responsive	Responsive	_____

High End Multichannel Monitor – 25 nos.

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1.	M/s. M M Medicals Systems.	BSM 3763	Responsive	Responsive	_____
2.	M/s. Trans Health Care India Pvt. Ltd.	General Meditech inc., China / G3L	Non Responsive	Responsive	As per the specialist report the model is not meeting the following points: 1. The weight of the Etco2 transducer at the patient end should be less than 25 gm. Should be able to

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
					<p>replace and supply EtcO2 transducer of less than 10 gms for NICU use if the end user wants it.</p> <p>2. Should have facility to detect sudden fall in BP in between periodic NIBP measurement to ensure precise continuous monitoring of BP.</p>
3.	M/s. Solokrafts Industries.	Mindray Benevision N15	Non Responsive	Responsive	<p>As per the specialist report the model is not meeting the following points:</p> <p>1. The weight of the EtcO2 transducer at the patient end should be less than 25 gm. Should be able to replace and supply EtcO2 transducer of less than 10 gms for NICU use if the end user wants it.</p> <p>2. Should have facility to detect sudden fall in BP in between periodic NIBP measurement to ensure precise continuous monitoring of BP.</p>
4.	M/s. Sri Sai Ekdant Surgicals.	Philips MX 550	Non Responsive	Responsive	<p>As per the specialist report the model is not meeting the following points:</p> <p>1. The weight of the EtcO2 transducer at the patient end should be less than 25 gm. Should be able to replace and supply EtcO2 transducer of less than 10 gms for NICU use if the end user wants it.</p> <p>2. Should have facility to detect sudden fall in BP in between periodic NIBP measurement to ensure precise continuous monitoring of BP.</p>