

07.09.2021

TENDER FOR FIXING RATE CONTRACT FOR SUPPLY OF CPAP MASK,HME FILTER, BACTERIAL/VIRAL FILTER, NASAL CANNULA, NON-REBREATHER MASK, DISPOSABLE VENTURI MASK, DISPOSABLE NEBULIZER KIT, DISPOSABLE OXYGEN MASK, DISPOSABLE VENTILATOR CIRCUIT AND OXYGEN HOOD

TENDER NO. 690/RC/TNMSC/ENGG/2021, DT:30.06.2021

Result of Evaluation

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
1.	a) CPAP Mask – Vented b) CPAP Mask – Non - Vented	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Non - Responsive	Responsive	As per specialist remarks: i. the mask frame does not have proper seal. ii. the headgear has no proper fitting and the Velcro strap is not sticking properly. Hence not recommended.
			Option B: Draeger	Responsive	Responsive	-----
		2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Non - Responsive	Responsive	As per specialist remarks: i. the mask frame does not have proper seal and the silicon cushion is not fitting in the frame if removed. ii.) headgear does not have proper fit.
		3. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: i. the silicon cushion of the mask is not fitting in the frame if removed. ii.) headgear does not have proper fit.
4. M/s. Dream Big Networks	Make: Amit Surgicals	Non - Responsive	Responsive	As per specialist remarks: i. the silicon cushion of the mask is not fitting in the frame if removed.		

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
						ii.) headgear does not have proper fit.
		5. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: i. the silicon cushion of the mask is not fitting in the frame if removed. ii.) headgear does not have proper fit.
		6. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: i. the silicon cushion of the mask is not fitting in the frame if removed. ii.) headgear does not have proper fit.
		7. M/s. Madras Healthcare	Make: Synergy Intact Pvt. Ltd., India	Non - Responsive	Responsive	As per specialist remarks: i. Poor quality of the mask and no proper fit. ii.) headgear does not have proper fit.
		8. M/s. Resmed India Pvt. Ltd.	Make: Curative Medical Technology Inc.	Responsive	Responsive	-----
		9. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		10. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		11. M/s. Medisafe International	Make: Medisafe International	Non - Responsive	Responsive	As per specialist remarks: i. cushion and frame not fixing properly. ii.) headgear does not have proper fit and the Velcro strap is of poor quality.
2.		1. M/s. Madras	Option A: JK Medical	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
	a) HME Filter - Adult	Surgicals and Chemicals	Option B: Draeger	Responsive	Responsive	-----
	b) HME Filter - Paediatric	2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
		3. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Responsive	Responsive	-----
		4. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----
		5. M/s. Vive Solutions	Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: Filter material not good. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.
		6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Responsive	Non - Responsive	Does not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		7. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: Only sponge is placed as filter.

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		8. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Responsive	Responsive	-----
		9. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Responsive	Responsive	-----
		10. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		11. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----
		12. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
		3.	a) Bacterial/ Viral Filter - Adult b) Bacterial/ Viral Filter - Paediatric as per specification	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Responsive
			Option B: Draeger	Responsive	Responsive	-----
		2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
		3. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Responsive	Responsive	-----
		4. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		5. M/s. Vive Solutions	Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: Filter material not good. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.
		6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Responsive	Non - Responsive	Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		7. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: Only sponge is placed as filter.
		8. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Responsive	Responsive	-----
		9. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Responsive	Responsive	-----
		10. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		11. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		12. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
4.	a) Nasal Cannula - Adult	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Responsive	Responsive	-----
			Option B: Draeger	Responsive	Responsive	-----
	b) Nasal Cannula - Paediatric	2. M/s. Morrisons Lifecare Pvt. Ltd.	Make: Morrisons Lifecare Pvt. Ltd.	Responsive	Responsive	-----
	c) Nasal Cannula - Neonatal	3. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
		4. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: Nasal tip very fragile and easily detachable.
		5. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		6. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----
		7. M/s. Vive Solutions	Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: Quality of the tube is not good. Kinking of tube. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
						of quoted items not furnished.
		8. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Responsive	Non - Responsive	Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		9. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: Quality of the tube is not good. Kinking of tube.
		10. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Responsive	Responsive	-----
		11. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		12. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
5.	a) Non - Rebreather Mask - Adult	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Responsive	Responsive	-----
			Option B: Draeger	Responsive	Responsive	-----
	b) Non - Rebreather Mask - Paediatric	2. M/s. Morrisons Lifecare Pvt. Ltd.	Make: Morrisons Lifecare Pvt. Ltd.	Responsive	Responsive	-----
			3. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive
	4. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: Quality of material is not good.	
	5. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----	
	6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: Quality of material is not good.	
	7. M/s. MC Dalal & Co.	Make: Life-O-Line Technologies, India	Non - Responsive	Responsive	As per specialist remarks: Quality of material is not good. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.	
8. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Non - Responsive	Responsive	As per specialist remarks: i. Quality of material is not good. ii. low resistance check valve on either side of the mask to prevent re-breathing through the mask is not available.		

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		9. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		10. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		11. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----
		12. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
6.	a) Disposable Oxygen Mask - Adult	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Responsive	Responsive	-----
			Option B: Draeger	Responsive	Responsive	-----
	b) Disposable Oxygen Mask - Paediatric	2. M/s. Morrisons Lifecare Pvt. Ltd.	Make: Morrisons Lifecare Pvt. Ltd.	Responsive	Responsive	-----
		3. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
		4. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: Quality of material is not good.
		5. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		6. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		7. M/s. Vive Solutions	Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: Quality of material is not good. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.
		8. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: Quality of material is not good. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		9. M/s. MC Dalal & Co.	Make: Life-O-Line Technologist, India	Non - Responsive	Responsive	As per specialist remarks: Quality of material is not good.
		10. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Non - Responsive	Responsive	As per specialist remarks: Quality of material is not good.
		11. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		12. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		13. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
7.	a) Disposable Venturi Mask - Adult	1. M/s. Madras Surgicals and Chemicals	Option A: JK Medical	Responsive	Responsive	-----
			Option B: Draeger	Responsive	Responsive	-----
	b) Disposable Venturi Mask - Paediatric	2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
			Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: i. Quality of mask not satisfactory ii. Only two adapters available and no standard colour coding provided.
		4. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----
			Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: i. Quality of product not satisfactory. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: i. Quality of mask not satisfactory ii. no standard colour coding provided for adapters and adapters for various O2 levels not provided. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		7. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Responsive	Responsive	-----
		8. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		9. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----
		10. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		11. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
		12. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: i. Quality of product not satisfactory
8.	a) Disposal Nebulizer kit – Adult	1. M/s. Madras Surgicals	Make: JK Medical	Responsive	Responsive	-----

S c h .	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
	b) Disposal Nebulizer kit – Paediatric	and Chemicals				
		2. M/s. Morrisons Lifecare Pvt. Ltd.	Make: Morrisons Lifecare Pvt. Ltd.	Responsive	Responsive	-----
		3. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
		4. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: i. Quality of mask not satisfactory. ii. the elastic head strap is of poor quality.
		5. M/s. Dream Big Networks	Make: Amit surgicals	Responsive	Responsive	-----
		6. M/s. Vive Solutions	Make: Oxycare Medical Devices	Non - Responsive	Non - Responsive	As per specialist remarks: i. Quality of mask not satisfactory. ii. the elastic head strap is of poor quality. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		7. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: i. Quality of mask not satisfactory. ii. the elastic head strap is of poor quality. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		8. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----
		9. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Responsive	Responsive	-----
		10. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	ISO/CE/USFDA certification not furnished
		11. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		12. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
		13. M/s. MC Dalal & Co.	Make: Life-O-Line Technologists, India	Non - Responsive	Responsive	As per specialist remarks: i. Quality of mask not satisfactory. ii. the elastic head strap is of poor quality.
9.	a) Disposable	1. M/s. Madras	Option A: JK Medical	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
.	Ventilator Circuit with double water trap- Neonatal	Surgicals and Chemicals	Option B: Draeger	Responsive	Responsive	-----
		2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
	Disposable Ventilator Circuit with double water trap – Paediatric	3. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good.
		4. M/s. Dream Big Networks	Make: Amit surgicals	Non - Responsive	Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good.
		5. M/s. Vive Solutions	Make: Oxycare Medical Devices, India	Non - Responsive	Non - Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.
		6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		7. M/s. Madras Healthcare	Make: Lakshmi Life Sciences Limited	Non - Responsive	Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good.
		8. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Responsive	Responsive	-----
		9. M/s. St. Luke Medical Devices	Make: St. Luke Med Flair, India	Responsive	Responsive	-----
		10. M/s. Medisafe International	Make: Medisafe International	Non - Responsive	Responsive	As per specialist remarks: i. quality of tube not good. ii. water trap quality not good.
		11. M/s. Flexicare (Group) Limited	Make: Flexicare	Responsive	Responsive	-----
10.	a) Oxygen Hood - Large	1. M/s. Madras Surgicals and Chemicals	Make: Nice Neotech	Non - Responsive	Responsive	As per specialist remarks: the oxygen nozzle is not compatible with universal oxygen tube.
	b) Oxygen Hood - Medium	2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Responsive	Responsive	-----
	c) Oxygen Hood - Small	3. M/s. Dream Big Networks	Make: Amit surgicals, India	Non - Responsive	Responsive	As per specialist remarks: The sample produced has sharp edges.

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
		4. M/s. Vive Solutions	Make: APHORB	Non - Responsive	Non - Responsive	As per specialist remarks: The sample produced was square shape and has sharp edges. Does not meet the following qualification criteria: i. affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence. ii. manufacturing capacity. iii. performance statement of the bidder for the supply of quoted items not furnished.
		5. M/s. Madras Healthcare	Make: Ibis Medical	Responsive	Responsive	-----
		6. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Responsive	Non - Responsive	Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
		7. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Responsive	Non - Responsive	Manufacturing capacity of Oxygen hood not furnished. Performance statement for Oxygen hood supply not furnished.
		8. M/s. Medisafe International	Make: Medisafe International	Responsive	Responsive	-----
1 1.	a) Disposable Ventilator Circuit with dual heated	1. M/s. Madras Surgicals and Chemicals	Make: Draeger	Responsive	Responsive	-----

S c h	Name of Item	Name of Bidders	Make	Responsiveness		Reason for rejection / Remarks
				Technical Specification	Commercial points	
	wire - Neonatal b) Disposable Ventilator	2. M/s. Lub Dub Medical Technologies Pvt. Ltd.	Make: Lub Dub	Non - Responsive	Responsive	As per specialist remarks: the patient circuit is not able to withstand the heat.
	Circuit with dual heated wire - Paediatric c) Disposable Ventilator	3. M/s. Solokrafts Industries	Make: Explore Synergy Synocare Pvt. Ltd., India	Non - Responsive	Non - Responsive	As per specialist remarks: Quality of the circuit not good. Did not meet the following qualification criteria: i. manufacturing capacity. ii. performance statement of the bidder for the supply of quoted items not furnished.
	Circuit with dual heated wire - Adult	4. M/s. Medifocus India Pvt. Ltd.	Make: Atrium Medical Technologies	Non - Responsive	Responsive	As per specialist remarks: Material of the circuit not compatible.
		5. M/s. Eurasia Healthcare	Make: Venticare Medical Inc.	Non - Responsive	Responsive	As per specialist remarks: Quality of the circuit not good.

Sd/-
General Manager(E)