

07.09.2021

**LOCAL COMPETITIVE BIDDING FOR SUPPLY AND INSTALLATION OF PFN
SET FOR MEDICAL COLLEGE HOSPITALS AT VARIOUS DISTRICTS OF
TAMIL NADU**

TENDER NO. PKG35(R)/LCB/TNUHP/TNMSC/ENGG/2021, DT:19.07.2021

RESULT OF EVALUATION

Sl. No	Name of Bidders	Make	Responsiveness		Reason for rejection
			Technical specification	Commercial points	
1.	M/s. Nirupama Medical Devices	Make: Changzhou Kanchui Medical Innovation Co. Ltd. (A Medtronic Company)	Responsive	Responsive	-----
2.	M/s. Esquire Surgicals	Make: S.H. Pitkar Ortho tools Pvt. Ltd., India	Non - Responsive	Responsive	Based on sample demonstration , as per Specialist remarks, the quoted model does not meet the following specification: 1. Spiral Blade 75 - 120mm in 5mm increment. One blade per nail along with a guide wire for each blade. - The locking mechanism is not proper - Proximal targeting device is not provided with compression option.

Sd/-

General Manager