

**RESULT OF EVALUATION**

**TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF  
BLOOD BANK REFRIGERATOR, BLOOD COLLECTION MONITOR, DONOR  
BLOOD COLLECTION COUCH (ELECTRONIC), CRYOBATH AND COLUMN  
AIDED GEL AGGLUTINATION TECHNOLOGY APPARATUS  
Tender No. 646/BBK/RC/TNMSC/ENGG/2021, Dt.01.02.2021**

**1. Blood Bank Refrigerator**

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Sun Biomedical Private Limited	Terumo Penpol / BR-160L	Responsive	Responsive	-----
2	M/s. Asean Scientific Express	Own/AS-180 BBR	Responsive	Responsive	-----
3	M/s. Mithra Enterprises	BR 180 /Remi	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
4	M/s.Solokraft s Industries	EBR-300/ ELTEK	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
5	M/s.Cryoscien tific systems Pvt. Limited,	Own /BBR-160	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.

**2. Donor Blood Collection Couch (Electronic)**

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Sun Biomedical	LC300/ Terumop	Responsive	Responsive	-----

Sl. No	Bidders Quoted	Model/ Make	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
	Private Limited	enpol			
2	M/s. Mithra Enterprises,	Desire 3X/Remi	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
3	M/s.Solokrafts Industries	EDC-2/ ELTEK	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
4	M/s.Span Healthcare Private Limited	Comfy couch-D4 / Qualimed	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not meeting the points below: 1.Dimensions: Length of the back rest:900 plus or minus 20mm Leg rest length:650 plus or minus 20 mm Length of arm rest:600 plus or minus 20 mm Width of arm rest:150 plus or minus 10 mm Overall length(trendelenburg position):2200 plus or minus 20 mm Overall width(both armrests fully open):1850 plus or minus 20 mm Overall height(Lifting column fully extended):1580 plus or minus 20 mm.

### 3. Blood collection monitor

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Sun Biomedical Private Limited	D601/Terumopenpol	Responsive	Responsive	-----
2	M/s.Span Healthcare Private Limited	Insignia/ BM-35X	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not meeting the points below: Ingress of water-IPXI, protected against dripping water with IPXI certificate.
3	M/s. Mithra Enterprises	BR10 Ultra/Remi	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
4	M/s. Solokrafts Industries	BCM-02/ ELTEK	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.
5	M/s.Pyrotech Electronics Pvt. Ltd	PY-BCM/ Pyrotech	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the specialist report the model was not demonstrated.

### 4. Cryobath

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Sun Biomedical Private Limited	Terumo Penpol & CB100	Responsive	Responsive	-----
2	M/s. Asean Scientific Express	Garuda PTB -20	Non Responsive	Not considered due to non responsive in technical	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
				specification	
3	M/s. Mithra Enterprises,	Remi & CB 704	Non Responsive	Not considered due to non responsive in technical specification	<b>Technical:</b> As per the report the specialist, the model is not demonstrated.

#### 5. Column Aided Gel Agglutination Technology Apparatus

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Madras Healthcare	Ortho Clinical Diagnostics	Responsive	Responsive	-----
2	M/s. Span Healthcare Private Limited	Spin dot & Therm dot /DOT Diagnostics	Responsive	Responsive	-----
3	M/s.AR Pharma & Distributors	Matrix Gel system/ CC0800	Responsive	Responsive	-----
		Matrix Gel system/ CW 2400			