

**Online Tender for Fixing Rate Contract for Supply and Installation of
Portable Echo Color Doppler Ultrasound and Color Doppler Ultrasound
Online tender ref: E952/USG/RC/TNMSC/ENGG/2023, Dt.30.03.2023**

EVALUATION REPORT

Sch(1) Portable Echo Color Doppler Ultrasound:

Sl. No.	Name of the Bidders Model/ Make	Responsiveness of Bidders		Reason of rejection by the specialist committee
		Technical Evaluation	Commercial Evaluation	
1	M/s. Esaote India (NS) Limited., Model: MyLab Omega, Make: Esaote S.p.A Italy	Responsive	Responsive	-----
2	M/s.Gemini Medical Networks Pvt.Ltd., Model: CHISON SONOBOOK8, Make: Chison Medical Technologies Co.Ltd.,	Responsive	Responsive	-----
3	M/s.Solokraft Industries., Model: M6, Make: Mindray	Responsive	Responsive	-----
4	M/s. Salvator Healthcare Pvt.Ltd., Model: SonoRad K40, Make: Trivitron	Non-Responsive	Non – Responsive	<p><u>Technical Evaluation:</u> As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended.</p> <ol style="list-style-type: none"> The system must have the ability to function by AC/DC or battery power with the same degree of functionality, the battery life (run time) shall be at least 2 (Two hours), this need to be demonstrated – Not Complied. B/W Thermal Printer – Not demonstrated. <p><u>Commercial Evaluation:</u> The bidder did not furnish the Purchase Order, Installation Report and performance report to meet the past qualification criteria of manufacturer and non-manufacturer bidder. Hence not recommended.</p>

Sch(2) Color Doppler Ultrasound Unit (Type I):

Sl. No.	Name of the Bidders Model/ Make	Responsiveness of Bidders		Reason of rejection by the specialist committee
		Technical Evaluation	Commercial Evaluation	
1	M/s.Gemini Medical Networks Pvt.Ltd., Model: CHISON Q Bit7, Make: Chison Medical Technologies Co.Ltd.,	Non-Responsive	Responsive	Technical Evaluation: As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended. 1. The Ultrasound scanner should have quick boot up time and able to deliver quality scanning within 30 seconds of boot up – Not Complied (63 Seconds).
2	M/s.Solokraft Industries., Model: DC30 FHD, Make: Mindray	Responsive	Responsive	-----
3	M/s. Salvator Healthcare Pvt.Ltd., Model: SonoRad V20, Make: Trivitron	Non-Responsive	Non-Responsive	Technical Evaluation: As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended. 1. The Ultrasound scanner should have quick boot up time and able to deliver quality scanning within 30 seconds of boot up – Not Complied (108 Seconds). Commercial Evaluation: The bidder did not furnish the Purchase Order, Installation Report and performance report to meet the past qualification criteria of manufacturer and non-manufacturer bidder. Hence not recommended.

Sch(3) Color Doppler Ultrasound Unit (Type II)

Sl.No	Name of the Bidders Model/ Make	Responsiveness of Bidders		Responsiveness of Bidders
		Technical Evaluation	Commercial Evaluation	
1	M/s. Esaote India (NS) Limited., Model: MyLab Omega, Make: Esaote S.p.A Italy	Responsive	Responsive	-----
2	M/s.Gemini Medical Networks Pvt.Ltd., Model: CHISON SONOBOOK8, Make: Chison Medical Technologies Co.Ltd.,	Non-Responsive	Responsive	Technical Evaluation: As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended.

SI.No	Name of the Bidders Model/ Make	Responsiveness of Bidders		Responsiveness of Bidders
		Technical Evaluation	Commercial Evaluation	
				1. The Ultrasound scanner should have quick boot up time and able to deliver quality scanning within 30seconds of boot up – Not complied (57 Sec)
3	M/s.Solokraft Industries., Model: M6, Make: Mindray	Responsive	Responsive	-----
4	M/s. Salvator Healthcare Pvt.Ltd., Model: SonoRad K40, Make: Trivitron	Responsive	Non-Responsive	Commercial Evaluation: The bidder did not furnish the Purchase Order, Installation Report and performance report to meet the past qualification criteria of manufacturer and non-manufacturer bidder. Hence not recommended.

Sch(4) Portable Ultrasound with Color Doppler System with standard probe:

SI. No.	Name of the Bidders Model/ Make	Responsiveness of Bidders		Responsiveness of Bidders
		Technical Evaluation	Commercial Evaluation	
1.	M/s. Esaote India (NS) Limited., Model: MyLab Omega, Make: Esaote S.p.A Italy	Responsive	Responsive	
2.	M/s.Solokraft Industries., Model: MX7, Make: Mindray	Non - Responsive	Responsive	As per the technical evaluation report furnished by the specialist, the model quoted by the bidder is not meeting the following point. Hence not recommended. 1. Laparoscopic Probe band width Frequency (4-13 Mhz) for Laparoscopic imaging application – Not complied. 2. Centre 0-degree Biopsy Convex array probe band width Frequency (1-8 Mhz) for General Imaging Application with reusable Biopsy Guide pinless probe connector – Not complied.

**Sd/-
General Manager (E)**