

22.06.2024

Draft Evaluation Result

Tender ref. no. E933/NIV/RC/TNMSC/ENGG/2023, Dt.15.03.2023

Online tender for fixing rate contract for supply and installation of CPAP machine with Bubble CPAP Kit and BIPAP for various Govt. Institutions

Sl. No	Name of the Bidder	Technical Evaluation Report	Commercial Evaluation Report	Reason for rejection
Sch.1 CPAP Machine with Bubble CPAP Kit				
1	M/s. Phoenix Medical Systems Private Limited Make: Phoenix Medical Systems Pvt. Ltd. Model: SVAAS	Responsive	Responsive	Nil
Sch.2 BIPAP				
1	M/s.ResMed India Pvt. Ltd. Make: ResMed Model: Stellar100	Responsive	Responsive	Nil
2	M/s. Hitwin Healthcare Private Limited Make: Breas Medical AB ("Breas") Model: Vivo 3	Non-Responsive	Responsive	As per the technical evaluation report of the Specialist and BME committee, the model quoted by the bidder did not meet the following tender specification points: 1. Integrated FiO2 not available. 2. Other deviations noted in the following tender specification: i. IPAP:4-40cm of water instead of 2-40cm ii. EPAP/PEEP 2-20 cmH2O instead of 2-25 iii. Respiratory rate: 4-40 BPM instead of 5-60 BPM Hence not recommended.