

Dt.29.07.2024

EVALUATION REPORT RESULT

Tender no.E1106/COT/RC/TNMSC/ENGG/2023, Dt.07.12.2023

**Online Tender for Fixing Rate contract for supply of Labour Cot with
Mattress to various Govt. Institutions in the State of Tamilnadu**

Sch.no.1 Labour Cot with Mattress

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Solokrafts Industries, Chennai Make: Solokrafts, Model: KRAFT 301	Responsive	Responsive	-----
2	M/s. Phoenix Medical Systems Pvt. Ltd, Chennai Make: Phoenix, Model: DT100	Responsive	Responsive	-----
3	M/s. Confident Dental Equipments Pvt. Ltd, Bangalore Make: Confident, Model: GYNB-054	Responsive	Responsive	-----
4	M/s. Madras Surgicals& Chemicals, Chennai Make: Imedfurns, Model: IMED 4015	Responsive	Responsive	-----
5	M/s. Godrej and Boyce Manufacturing Company Limited, Mumbai Make: Godrej, Model: Solace Lite-Apo	Non-Responsive	Responsive	The quoted model not meeting the following technical points:- <ul style="list-style-type: none">• The bed should be suitable for use during Delivery / scan purpose, (3 Part) Retracting leg section and two section mattress allowing maximum access to mother & baby during delivery – <u>Mattress is in 2 sections only.</u>• Should be supplied with one Squat bar (S.S.) & one baby receiving tray – <u>Not shown during demonstration.</u>• Should have paper holder with roller. Should have at least 12

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				<p>L capacity fluid basin made of polymer - <u>Not shown during demonstration.</u></p> <ul style="list-style-type: none"> The quoted model is very wide and difficult to access patient by 2 persons on either side at same time. Hence not recommended.
6	<p>M/s. South India Surgical Co. Ltd, Chennai Make: Bio Breeze, Model: HS-105</p>	Non-Responsive	Non-Responsive	<ol style="list-style-type: none"> Not demonstrated Did not furnish the enduser certificate along with purchase order and installation report to meet the qualification criteria.
7	<p>M/s. Samar Steel Industries, Maharashtra Make: Samar, Model: SSI-HF-101</p>	Non-Responsive	Non-Responsive	<ol style="list-style-type: none"> Not demonstrated Did not furnish the enduser certificate along with purchase order and installation report to meet the qualification criteria.
8	<p>M/s. SS Groups, Chennai Make: Meditek, Model: 5301</p>	<p><u>Bid security not furnished.</u></p> <p>"Domestic SS Enterprises who are the manufacturers of quoted product within the State and registered with MSME are alone eligible for EMD exemption, otherwise not eligible". Hence their bid is treated as non-responsive.</p>		

Sd/-
General Manager (E)