

Dt.06.08.2024

EVALUATION REPORT RESULT

Tender no.E1133/EYE/RC/TNMSC/ENGG/2024, Dt.07.02.2024

Online Tender for Fixing Rate Contract for Supply and Installation of Slit Lamp with Refraction OPD Chair Unit, Slit Lamp with camera, OPD Chair unit and Slit Lamp Basic to various Govt. Institutions in the State of Tamilnadu

Sch.no.1 Slit Lamp with Refraction OPD Chair Unit

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: AIA 11 3SL with AARU 2000 Elegant	Responsive	Responsive	-----
2	M/s. Aurolab Trust, Make: Labotech, Model: Hawki - T3 Led OPD CU	Responsive	Responsive	-----

Sch.no.2 Slit Lamp with camera

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: SL-CAM ICM-7	Responsive	Responsive	-----
2	M/s. Aurolab Trust, Make: Labotech, Model: Hawki – T5 Led	Responsive	Responsive	-----
3	M/s. Remidio Innovative Solutions Pvt. Ltd, Make: Remidio, Model: EVO-45	Non-Responsive	Responsive	The quoted model not meeting the following technical points:- <ul style="list-style-type: none"> • Image resolution: 2592x1944 – Not shown during demonstration. • Video resolution: 2592x1944 – Not shown during demonstration. • Frame/sec: 23 fps at 648x486 pixel, 15fps at 1296x972 pixel, 6fps at 2592x1944 pixel - Not shown during demonstration. • Live zoom – Not available.

Sch.no.3 OPD Chair unit

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: AARU 2000 Elegant	Responsive	Responsive	-----
2	M/s. Aurolab Trust, Make: Two-M, Model: OPD CU 2MM Chair unit	Responsive	Responsive	-----

Sch.no.4 Slit Lamp Basic

Sl. No	Name of the Bidder	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Appasamy Associates Pvt. Ltd, Make: Appasamy, Model: AIA 11 3SL	Responsive	Responsive	-----
2	M/s. Aurolab Trust, Make: Labotech, Model: Hawki - T3 Halogen (Basic)	Responsive	Responsive	-----
3	M/s. Remidio Innovative Solutions Pvt. Ltd, Make: Remidio, Model: SLX-45	Responsive	Responsive	-----
4	M/s. Alan Medical & Laboratory Products, Make: Kwaliti Micro Scientific, Model: Series 3S	Non-Responsive	Responsive	The quoted model not meeting the following technical points:- <ul style="list-style-type: none"> • Slit length: more than 52-78 mm – Not available. • Slit apertures from 0-12 mm or more continuous or in 6 steps – Not available.

**Sd/-
General Manager (E)**