

RESULT OF RE-EVALUATION OF SHORT LISTED BIDDERS

ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION OF ELISA READER, ELISA WASHER AND DI-ELECTRIC TUBE SEALER

Tender No. E1120/ELISA/RC/TNMSC/ENGG/2024, DT: 02.01.2024

Sch. No.1. Elisa Reader (without Accessories)

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Shan biotech & Diagnostics	Robonik/ Readwell Touch	Responsive	Responsive	-----
2	M/s. Transasia Bio Medicals Ltd	Erba/Lisa Scan EM	Responsive	Responsive	-----
3	M/s. Meril Diagnostics Pvt Ltd	Meril/ Merilyzer EIA Quant	Responsive	Responsive	-----
4	M/s. Sun bio Medicals Private Limited	Tulip/ Lisequant/ IT	Responsive	Responsive	-----

Sch. No.2. Elisa Washer

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Shan biotech & Diagnostics	Robonik/ Washwell Plate	Responsive	Responsive	-----
2	M/s. Transasia Bio Medicals Ltd	Erba/Lisa wash	Responsive	Responsive	-----
3	M/s. Meril Diagnostics Pvt Ltd	Meril/ Merilyzer EIA Quant	Non Responsive	Responsive	As per specialist report the model is not meeting the below points: 1. Dispensable volume up to 1000µl. Soaking time 1-600sec. 2. Should be built-in for minimum of one microplate.

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
4	M/s. Sun bio Medicals Private Limited	Tulip/ Lisawash 4000	Responsive	Responsive	-----

Sch. No.3. Di-Electric Tube Sealer

Sl. No	Bidders Quoted	Make/ Model	Responsiveness		Reason for rejection / Remarks
			Technical evaluation report	Commercial evaluation report	
1	M/s. Sun bio Medicals Private Limited	Terumo Penpol/ LM100E	Responsive	Substantially Responsive	-----