

Dt.31.08.2024

EVALUATION REPORT RESULT**Tender no.E1053/EYE/DME/TNMSC/ENGG/2023, Dt.13.09.2023**

Online Tender for Fixing Rate Contract for Supply and Installation of Posterior Segment Operating Microscope with Assistant Scope Camera Monitor and BIOM, Pars Plana Vitrectomy Machine – High Speed and Anterior Vitrectomy Unit to Regional Institute of Ophthalmology, Chennai

Sl. No	Name of the equipment and bidder with make and model	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
Sch.no.1 Posterior Segment Operating Microscope with Assistant Scope Camera Monitor and BIOM – 1 no.				
1	M/s. Asian Vision, Chennai Make: Takagi, Japan, Model: OM-9	Responsive	Responsive	-----
Sch.no.2 Pars Plana Vitrectomy Machine – High Speed – 1 no.				
1	M/s. Toshbro Medicals Pvt. Ltd, Mumbai Make: Oertli Instruments AG, Switzerland, Model: OS4	Responsive	Responsive	-----
2	M/s. KK Allianza, Bangalore Make: Optikon 2000 S.p.A, Italy, Model: R-EVOLUTION	Responsive	Non-Responsive	Did not furnish the enduser certificate along PO & installation report to meet the qualification criteria. The document furnished is different make (Labotech) against the quoted make (Optikon).
3	M/s. Nithura Medical Systems, Chennai Make: Oscar Vision Tech, Chennai, Model: Micro ViT AB6K	Non-Responsive	Non-Responsive	1) The quoted model did not meeting following technical points:- <ul style="list-style-type: none"> • Endolaser – inbuilt embedded 532 nm endolaser controller from the main panel. Shall have the power source for LIO on console itself – <u>No inbuilt laser source.</u> • Powerful intraocular illumination by both hand

Sl. No	Name of the equipment and bidder with make and model	Responsiveness of bidders		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
				held and chandelier options. Endo illuminator & Chandelier – <u>Chandelier option not available.</u> 2) Did not furnish the manufacturing capacity and enduser certificate along PO & installation report to meet the qualification criteria.
Sch.no.3 Anterior Vitrectomy Unit – 4 nos.				
1	M/s. Nithura Medical Systems, Chennai Make: Oscar Vision Tech, Chennai, Model: AB60	Responsive	Responsive	-----

**Sd/-
General Manager (E)**