

Dt.26.09.2024

EVALUATION REPORT RESULT

Online quotation for supply and installation of High End Surgical Suction Pump to Dept. of Neurosurgery, Govt. Kilpauk Medical College Hospital, Chennai

Quotation no.E769/LT/SUC/NHM/TNMSC/ENGG/2024, Dt.16.02.2024

Sl. No	Name of the bidder	Status	Reason for Rejection
1	M/s. Miti Electronics Pvt. Ltd, Make: Medela, Switzerland Model: Dominant Flex	Recommended	-----
2	M/s. RL Healthcare, Make: ElmaslarMedikal, Model: Lifetime SA01HT W/Sensor Hospital Type Surgical Aspirator	Not-Recommended	The quoted model does not have CF electrical certification for providing protection against electrical shock. Hence not recommended.
3	M/s. The Advanced Medical Systems Inc, Make: Atmos, Germany Model: Record 55 DDS	Not-Recommended	1) The following points are not mentioned in the technical compliance statement:- <ul style="list-style-type: none">• Should provide heat resistant glass free piston cylinder with flat spring technology.• Should supply with membrane vacuum regulator.• Should be available safety jars with over flow protection device.• Should supply with standard rail which is attached in the mobile trolley.• System should have type CF electrical certification for providing protection against electrical shock.• System should be complaint with IP 21 Ingress protection to prevent ingress of foreign bodies and any type of fluid inside the machine.• System should be oil free and should not contain any glass or stainless steel components.• System should have colour coded LED

Sl. No	Name of the bidder	Status	Reason for Rejection
			<p>indicators for status of the machine standby, running and error.</p> <ul style="list-style-type: none"> • Should be provided with tissue collector – 20 nos. per unit. <p>2) Tubing of 5m length and made of silicon to be provided – offered 2m length.</p> <p>3) Noise levels should be between 38-40db – Upto 50 db. Hence not recommended.</p>

Sd/-
General Manager (E)(i/c)