

TENDER NO: E1140/MON/RC/TNMSC/ENGG/2024, DT: 07.02.2024

**ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND
INSTALLATION OF HIGH END MULTIPARA MONITOR TO VARIOUS GOVT.
INSTITUTIONS**

Result of Evaluation

High End Multipara Monitor Type I

Sl. No	Name of the Bidders with Model / Make	Responsiveness		Reason for rejection / Remarks
		Technical evaluation report	Commercial evaluation report	
1	M/s. Pixel Imaging Systems. Make: Philips Model: Intellivue MX550	Responsive	Non-Responsive	The Manufacturer M/s. Philips did not agree to furnish the undertaking as required to retain the bidder for 10 years
2	M/s. Solokrafts Industries. Make: Mindray, China Model: BeneVision N15 with BeneVision M1 transport Monitor	Non-Responsive	Responsive	Not meeting the following point: <ul style="list-style-type: none">• Modular system upgradeable to <u>Non-invasive hemoglobin (Hb) transcutaneous PO2 and PCO2</u>, Spirometry, ScVO2, minimal invasive cardiac output, EEG, AGM and Mainstream ETCO2 through interchangeable modules (prices to be quoted separately).• Upgradable to Pleth variability index (PVI) & non-invasive Hemoglobin display (numerical and/or graphical) (Price to be quoted separately) Hence not recommended.
3	M/s. Schiller Healthcare Make: Guangdong Biolight Meditech Co., Ltd., China Model: Truscope P15	Non-Responsive	Responsive	Not meeting the following point: <ul style="list-style-type: none">• Modular system upgradeable to <u>Non-invasive hemoglobin (Hb) transcutaneous PO2 and PCO2</u>, Spirometry, ScVO2, minimal invasive cardiac output, EEG, AGM and Mainstream ETCO2 through interchangeable modules (prices to be quoted separately).• Upgradable to Pleth variability index (PVI) & non-invasive Hemoglobin display (numerical and/or graphical)

				(Price to be quoted separately) Hence not recommended.
4	M/s. M M Medicals Systems. Make: Nihon Kohden, Japan Model: Lifescope VS G5	Non Responsive	Responsive	Not meeting the following point: <ul style="list-style-type: none"> Modular system upgradeable to <u>Non-invasive hemoglobin (Hb) transcutaneous PO2 and PCO2</u>, Spirometry, ScVO2, minimal invasive cardiac output, EEG, AGM and Mainstream ETCO2 through interchangeable modules (prices to be quoted separately). Upgradeable to Pleth variability index (PVI) & non-invasive Hemoglobin display (numerical and/or graphical) (Price to be quoted separately) Hence not recommended.

High End Multipara Monitor Type II

Sl. No	Name of the Bidders with Model / Make	Responsiveness		Reason for rejection / Remarks
		Technical evaluation report	Commercial evaluation report	
1	M/s. Solokrafts Industries. Make: Mindray, China Model: BeneVision N22 with BeneVision N1 transport Monitor	Responsive	Responsive	-----
2	M/s. Schiller Healthcare Make: Guangdong Biolight Meditech Co., Ltd., China Model: Truscope P22	Responsive	Responsive	-----
3	M/s. M M Medicals Systems. Make: Nihon Kohden, Japan Model: Lifescope VS G7	Substantially -responsive	Responsive	-----