

**ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND
INSTALLATION OF HEATED HUMIDIFIED HIGH FLOW NASAL CANNULA
DEVICE TO VARIOUS GOVT. INSTITUTIONS**

E1065/HFO/NHM/TNMSC/ENGG/2023, DT: 13.09.2023

Heated Humidified High Flow Nasal Cannula Device

Sl. No	Name of the Bidders with Model / Make	Responsiveness		Reason for rejection / Remarks
		Technical evaluation report	Commercial evaluation report	
1	M/s. Indian Medical Systems & Technologies Make: M/s. Fisher & Paykel, Model: Airvo 3	Responsive	Responsive	-----
2	M/s. Hild Lifesciences Pvt. Ltd. Make: Hild Lifesciences Pvt. Ltd., India Model: Alto flow	Non Responsive	Non Responsive	<p>As pre specialist demo report :-</p> <p>A. On deploying oxygen post, there is a continuous alarm to indicate oxygen disconnection. The alarm is not a sensor at the patient level O2 delivery and make it difficult for machine portability. No provision available (flow meter) for connecting oxygen cylinder.</p> <p>B. The product appears flimsy and not durable.</p> <p>C. Cannula not available (HFNC) for demo.</p> <p>D. Not available for Tracheostomized patient.</p> <p>Hence non-responsive.</p> <p>Despite given chance the bidder did not furnish documentary evidence meeting the following terms of the tender as manufacturer bidder:-</p> <p>A. The manufacturer should be in continuous business of manufacturing / supplying and after sale services of products similar to that specified in the 'Schedule of</p>

				<p>requirement' during the last 5 years prior to bid opening –</p> <p>Company was incorporated on 19.08.2021.</p> <p>B. The bidder should submit an affidavit confirming that the performance statement given is correct along with copy of purchase order, copy of invoices, proof of payment received from Purchasers, documentary evidence (end user certificate)</p> <p>C. Documentary evidence (end users certificate) in support of the satisfactory operation of the goods are less than one year.</p> <p>Hence non-responsive.</p>
3	<p>M/s. Clarity Bio-Med Services Make: Heka Medicals India Pvt. Ltd. Model: HM-P-500</p>	Non Responsive	Non Responsive	<p>As pre specialist demo report :-</p> <p>A. No Inbuilt battery</p> <p>B. No facility available for disinfection.</p> <p>C. No O2 holder.</p> <p>D. No Inbuilt nebulisation facility.</p> <p>Hence non-responsive.</p> <p>Despite given chance the bidder did not furnish documentary evidence meeting the following terms of the tender as a non-manufacturer bidder:-</p> <p>A. Documentary evidence (end users certificate) in support of the satisfactory operation of the goods are less than one year.</p> <p>B. Not agreed to tender delivery condition.</p> <p>C. Not agreed to tender warranty condition.</p> <p>D. Not confirmed to 10 years of spare parts.</p> <p>Hence non-responsive.</p>