

EVALUATION REPORT

TENDER NO. E1146/SCOPE/DME-11MCH/TNMSC/ENGG/24,

DT.08.02.2024

ONLINE TENDER FOR SUPPLY AND INSTALLATION OF ARTHROSCOPY COMPLETE SET TO 11 NEW MEDICAL COLLEGE HOSPITALS

SCH. 1 ARTHROSCOPY COMPLETE SET

Sl. No	Name of the Bidder	Responsiveness		Reason for rejection / Remarks
		Technical evaluation report	Commercial evaluation report	
1	M/s. Avana Medical Devices Pvt. Ltd.	Substantially Responsive	Responsive	The responsiveness is subject to supply the separate console without change in the quoted rate.
2	M/s. Cnergy Orthocare (P) Ltd.	Non Responsive	Non Responsive	<p>a. The battery operated drill and saw system, Burr and cutters could be operated only with 10000 RPM against the requirement of 12000 RPM.</p> <p>b. The trauma and Big Bone Drills and Saws cannot be operated with the console unit.</p> <p>c. Not demonstrated irrigation system for high speed drills.</p> <p>Commercial Evaluation</p> <p>a. Did not furnish the appropriate end user certificate to meet the past qualification criteria for both manufacturer and Non-manufacturer.</p>
3	M/s. Indian Orthocare Surgicals	Responsive	Responsive	-----

**Sd/-
General Manager (E)**