

Date: 01.12.2024

**ONLINE TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND  
INSTALLATION OF CTG MONITOR WITH WIRED PROBE AND CTG  
MONITOR WITH WIRELESS PROBE TO GOVT. MEDICAL COLLEGE  
INSTITUTIONS IN THE STATE OF TAMILNADU**

**E921/CTG/RC/TNMSC/ENGG/2023, DT: 22.02.2023**

**RESULT OF EVALUATION**

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
<b>CTG Monitor with wired probe (Sch.no.1)</b>				
1.	M/s RL Healthcare, Thoothukudi <b>Make:</b> Shenzhen Luckcome Technology Inc., Ltd., China <b>Model:</b> L8 Fetal Monitor	Responsive	<b>Non- Responsive</b>	<b>Commercial points:</b> 1) The following documents pertaining to commercial terms are not furnished by them: a) Importers Photocopy of license renewed up to date. b) Documentary evidence of Constitution of firm. c) Authorization of Senior responsible officer of the Company to transact business. d) Copies of Balance Sheet and Profit & Loss Account for three years certified by the Auditors. e) GST registration no. for supply and services. f) GST Clearance Certificate as on 31.03.2022.

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>g) Performance Statement – Section X.</p> <p>h) The annual sales turnover for the bidder should be at least (Rs.139 lakhs for SI.No.1 &amp; Rs.25 lakhs for SI.No.2) in any one of the last three years (2019-20, 2020-21 &amp; 2021-22) certified by chartered accountants.</p> <p>i) The Cash/Credit limit of (Rs.34.75 lakhs for SI.No.1 &amp; Rs.6.25 lakhs for SI.No.2) with their bankers exclusively for this contract furnished by their bankers.</p> <p>j) Affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence (End User Certificate) in support of satisfactory Completion of orders.</p> <p>k) <u>Manufacturer bidders:</u> The bidder must have manufactured and supplied satisfactorily the similar model quoted in each schedule of the Schedule of</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>Requirements either directly or through of any other authorized dealer to the extent of <u>at least 3 Nos. for Sl. No.1 &amp; 1 No. for Sl. No.2</u> in any one of the last five years of which minimum 1 No. for Sl. No.1 &amp; 2 should have been supplied to Indian Institution and should be in use satisfactorily with no adverse report for at least one year preceding the date of bid opening.</p> <p>l) <u>Non-Manufacturer Bidder:</u> The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of <u>at least 1 No. for Sl. No.1 &amp; 2</u> of the quantities of similar model in last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.</p> <p>m) Section XIII: Self declaration by the bidder</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				for not subject to debarment.  2) As per letter ref. E1142/ECT/DME-11MCH/TNMSC/ENGG/2024 Dt. 12.09.2024, this firm is blacklisted for a period of 1 year from the date of receipt of notification, in which it was stated that they are ineligible to be selected in any of the tenders in process and also to participate in future tenders during the blacklisted period of 1 year.
2.	M/s Solokrafts Industries, Chennai <b>Make:</b> Shenzhen Comen Medical Instruments Co. Ltd., China <b>Model:</b> Star5000E/C11	Responsive	Responsive	----
3.	M/s Miti Electronics, Chennai <b>Make:</b> Bistos Co. Ltd., Korea <b>Model:</b> BT-350	Responsive	Responsive	----
4.	M/s Focus Imaging Systems, Chennai	Responsive	Responsive	----

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
	<b>Make:</b> Edan Instruments Inc., China <b>Model:</b> F3			
5.	M/s Thirupathi Enterprises, Kovilpatti* <b>Make:</b> Bistos Co. Ltd., Korea <b>Model:</b> BT-350	<b>Non-Responsive</b>	<b>Non-Responsive</b>	<p><b>Technical point:</b> Not demonstrated the equipment.</p> <p><b>Commercial points:</b> 1) Despite request in email dt. 21.08.2024, the following documents not furnished: a) Performance Statement – Section X b) Manufacturing Capacity of at least (1134 Nos. for SI.No.1 &amp; 180 Nos. for SI.No.2) similar capacity machines per annum certified by licensing authority/ chartered accountant. c) Cash/Credit limit of (Rs.34.75 lakhs for SI.No.1 &amp; Rs.6.25 lakhs for SI.No.2) with their bankers exclusively for this contract and a certificate furnished by their bankers. d) Affidavit conforming that the performance statement given is correct along with</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>copy of purchase order, copy of invoices, documentary evidence (End User Certificate) in support of satisfactory Completion of orders and function.</p> <p>e) GST Clearance certificate as on 31.03.2022.</p> <p><u>f) Manufacturer Bidders:</u> The bidder must have manufactured and supplied satisfactorily the similar model quoted in each schedule of the Schedule of Requirements either directly or through of any other authorized dealer to the extent of <u>at least 3 Nos. for Sl. No.1 &amp; 1 No. for Sl. No.2</u> in any one of the last five years of which minimum 1 No. for Sl. No.1 &amp; 2 should have been supplied to Indian Institution and should be in use satisfactorily with no adverse report for at least one year preceding the date of bid opening.</p> <p><u>g) Non-Manufacturer Bidder:</u> The bidder, as authorized by the manufacturer, must</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>have supplied and provided after sales service to the extent of <u>at least 1 No. for Sl. No.1 &amp; 2</u> of the quantities of similar model in last five (5) years to Indian Institution, which must be in satisfactory operation at least for one year on the date of bid opening.</p> <p>h) Bid document signed at all pages furnished: Corrigendum only furnished. Bid document not furnished.</p> <p>i) Validity of Bid (90 days)</p> <p>j) Delivery Schedule as per bid document (30 days from the date of receipt of the PO)</p> <p>k) Payment Terms as per bid document.</p> <p>l) Warranty is as per bid document (36 months from the acceptance of goods).</p> <p>m) Spares parts and after sales service facilities in India for atleast a period of 10 years.</p> <p>n) Section XIII: Self declaration by the bidder</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				for not subject to debarment.  2) Further, in email dt. 21.11.2024, OEM M/s Bistos Co. Ltd., Korea had stated that they nullify the Manufacturer's Authorization furnished by them earlier.
6.	M/s Sabisu, Chennai <b>Make:</b> Bistos Co. Ltd., Korea <b>Model:</b> BT-350	Responsive	<b>Non-Responsive</b>	<b>Commercial point:</b> In email dt. 21.11.2024, OEM M/s Bistos Co. Ltd., Korea had stated that they nullify the Manufacturer's Authorization furnished by them earlier.
7.	M/s Star Medisystems, Chennai <b>Make:</b> Biolight Co. Ltd., China <b>Model:</b> F80	<b>Non-Responsive</b>	Responsive	<b>Technical point:</b> <i>As per the demo report furnished by the Specialists, the model quoted by them doesn't comply with the following technical specification: <u>Should be portable type weight less than 4kg.</u></i>
8.	M/s Maestros Electronics and Telecommunications Systems Limited, Navi Mumbai	<b>Non-Responsive</b>	Responsive	<b>Technical point:</b> <i>As per the demo report furnished by the Specialists, the model quoted by them doesn't comply with the</i>



S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
	<p><b>Make:</b> Maestros Electronics and Telecommunications Systems Limited, India</p> <p><b>Model:</b> CTG 150</p>			<p>following technical specification:  <u>Special high sensitive probe.</u></p>
9.	<p>M/s Surya Surgical Industries, Chennai</p> <p><b>Make:</b> BAMC Medical Limited, India</p> <p><b>Model:</b> Cardiac V (wired)</p>	Responsive	Responsive	----
<b>CTG Monitor with wireless probe (Sch.no.2)</b>				
1.	<p>M/s RL Healthcare, Thoothukudi</p> <p><b>Make:</b> Shenzhen Luckcome Technology Inc., Ltd., China</p> <p><b>Model:</b> L8ME Fetal Monitor</p>	Responsive	<b>Non-Responsive</b>	<p><b>Commercial points:</b></p> <p>1) The following documents pertaining to commercial terms are not furnished by them:</p> <p>a) Importers Photocopy of license renewed up to date.</p> <p>b) Documentary evidence of Constitution of firm.</p> <p>c) Authorization of Senior responsible officer of the Company to transact business.</p> <p>d) Copies of Balance Sheet and Profit &amp; Loss Account for three years certified by the Auditors.</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>e) GST registration no. for supply and services.</p> <p>f) GST Clearance Certificate as on 31.03.2022.</p> <p>g) Performance Statement – Section X.</p> <p>h) The annual sales turnover for the bidder should be at least (Rs.139 lakhs for SI.No.1 &amp; Rs.25 lakhs for SI.No.2) in any one of the last three years (2019-20, 2020-21 &amp; 2021-22) certified by chartered accountants.</p> <p>i) The Cash/Credit limit of (Rs.34.75 lakhs for SI.No.1 &amp; Rs.6.25 lakhs for SI.No.2) with their bankers exclusively for this contract furnished by their bankers.</p> <p>j) Affidavit conforming that the performance statement given is correct along with copy of purchase order, copy of invoices, documentary evidence (End User Certificate) in support of satisfactory Completion of orders.</p> <p>k) <u>Manufacturer bidders</u>:</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>The bidder must have manufactured and supplied satisfactorily the similar model quoted in each schedule of the Schedule of Requirements either directly or through of any other authorized dealer to the extent of <u>at least 3 Nos. for Sl. No.1 &amp; 1 No. for Sl. No.2</u> in any one of the last five years of which minimum 1 No. for Sl. No.1 &amp; 2 should have been supplied to Indian Institution and should be in use satisfactorily with no adverse report for at least one year preceding the date of bid opening.</p> <p>I) <u>Non-Manufacturer Bidder:</u> The bidder, as authorized by the manufacturer, must have supplied and provided after sales service to the extent of <u>at least 1 No. for Sl. No.1 &amp; 2</u> of the quantities of similar model in last five (5) years to Indian Institution, which must be in satisfactory operation at least for one</p>

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
				<p>year on the date of bid opening.</p> <p>m) Section XIII: Self declaration by the bidder for not subject to debarment.</p> <p>2) As per letter ref. E1142/ECT/DME-11MCH/TNMSC/ENGG/2024 Dt. 12.09.2024, this firm is blacklisted for a period of 1 year from the date of receipt of notification, in which it was stated that they are ineligible to be selected in any of the tenders in process and also to participate in future tenders during the blacklisted period of 1 year.</p>
2.	<p>M/s Focus Imaging Systems, Chennai</p> <p><b>Make:</b> Edan Instruments Inc., China</p> <p><b>Model:</b> F6 with Fetal telemetry system (FTS- 3)</p>	Responsive	Responsive	----
3.	M/s Miti Electronics, Chennai	Responsive	Responsive	----

S.No.	Equipment details	Responsiveness		Reasons for rejection
		Technical	Commercial	
	<p><b>Make:</b> Janitri Innovations Pvt. Ltd., India</p> <p><b>Model:</b> Keyar DT max</p>			
4.	<p>M/s Surya Surgical Industries, Chennai</p> <p><b>Make:</b> BAMC Medical Limited, India</p> <p><b>Model:</b> Cardiac V (wireless)</p>	Responsive	Responsive	----
5.	<p>M/s Maestros Electronics and Telecommunications Systems Limited, Navi Mumbai</p> <p><b>Make:</b> Maestros Electronics and Telecommunications Systems Limited, India</p> <p><b>Model:</b> CTG 150</p>	<b>Non-Responsive</b>	Responsive	<p><b>Technical point:</b></p> <p><i>As per the demo report furnished by the Specialists, the model quoted by them doesn't comply with the following technical specification:</i></p> <p><u>Special high sensitive probe.</u></p>

**Sd/-**  
**General Manager(E)(i/c)**