

23.12.2024

EVALUATION REPORT

TENDER NO. E1209/SCOPE/RC/TNMSC/ENGG/2024, DT. 07.08.2024

TENDER FOR FIXING RATE CONTRACT FOR SUPPLY AND INSTALLATION VIDEO
LARYNGOSCOPE TO GOVT. INSTITUTIONS

SCH. 1 VIDEO LARYNGOSCOPE

Sl. No	Name of the Bidder Make/Model	Responsiveness		Reason for rejection
		Technical evaluation report	Commercial evaluation report	
1	M/s. Care and Care Medical Systems Model : Allied Sharpscope 100 Make : Allied Medical	Responsive	Responsive	-----
2	M/s. Lakshmi Agency Model : S-VL-01 Make : Sensorendo Meditech Inc., China HOST (3502)	Non Responsive	Responsive	Technical Demonstration The quoted model does not fit for Neonatal Patients and also Neonatal blade not available, hence non recommended.

**Sd/-
General Manager (E)**